



## Full wwPDB EM Validation Report ⓘ

Mar 8, 2025 – 12:08 pm GMT

EMDB ID : EMD-19782  
Title : M. tuberculosis gyrase holocomplex with 150 bp DNA and BDM71403  
Authors : Gedeon, A.; Yab, E.; Dinut, A.; Sadowski, E.; Capton, E.; Dreneau, A.; Gioia, B.; Piveteau, C.; Djaout, K.; Lecat, E.; Wehenkel, A.M.; Gubellini, F.; Mechaly, A.; Alzari, P.M.; Deprez, B.; Baulard, A.; Aubry, A.; Willand, N.; Petrella, S.  
Deposited on : 2024-03-04  
Resolution : 2.80 Å(reported)

This is a Full wwPDB EM Validation Report for a publicly released PDB entry.

We welcome your comments at [validation@mail.wwpdb.org](mailto:validation@mail.wwpdb.org)

A user guide is available at

<https://www.wwpdb.org/validation/2017/EMMapValidationReportHelp>  
with specific help available everywhere you see the ⓘ symbol.

The types of validation reports are described at

<http://www.wwpdb.org/validation/2017/FAQs#types>.

---

The following versions of software and data (see [references ⓘ](#)) were used in the production of this report:

EMDB validation analysis : **FAILED**  
Validation Pipeline (wwPDB-VP) : 2.41

# 1 Experimental information

Property	Value	Source
EM reconstruction method	SINGLE PARTICLE	Depositor
Imposed symmetry	Not Provided	
Number of particles used	987000	Depositor
Resolution determination method	FSC 0.143 CUT-OFF	Depositor
CTF correction method	Not provided	
Microscope	FEI TITAN KRIOS	Depositor
Voltage (kV)	300	Depositor
Electron dose ( $e^-/\text{\AA}^2$ )	40.0	Depositor
Minimum defocus (nm)	0.9	Depositor
Maximum defocus (nm)	1.9000000000000001	Depositor
Magnification	Not provided	
Image detector	GATAN K3 BIOQUANTUM (6k x 4k)	Depositor