



## Full wwPDB EM Validation Report ⓘ

Jul 1, 2024 – 12:09 PM JST

EMDB ID : EMD-35019  
Title : MAVS fibrils surrounding the mitochondria  
Authors : Wang, P.; Guo, Q.  
Deposited on : 2022-12-21  
Resolution : Not provided

This is a Full wwPDB EM Validation Report for a publicly released PDB entry.

We welcome your comments at [validation@mail.wwpdb.org](mailto:validation@mail.wwpdb.org)

A user guide is available at

<https://www.wwpdb.org/validation/2017/EMTomogramValidationReportHelp>  
with specific help available everywhere you see the ⓘ symbol.

The types of validation reports are described at

<http://www.wwpdb.org/validation/2017/FAQs#types>.

---

The following versions of software and data (see [references ⓘ](#)) were used in the production of this report:

EMDB validation analysis : 0.0.1.dev92  
Validation Pipeline (wwPDB-VP) : 2.37.1


# 1 Experimental information

| Property                             | Value                       | Source    |
|--------------------------------------|-----------------------------|-----------|
| EM reconstruction method             | TOMOGRAPHY                  | Depositor |
| Imposed symmetry                     | Not Provided                |           |
| Number of tilted images used         | 60                          | Depositor |
| Resolution determination method      | Not provided                |           |
| CTF correction method                | Not provided                |           |
| Microscope                           | FEI TITAN KRIOS             | Depositor |
| Voltage (kV)                         | 300                         | Depositor |
| Electron dose ( $e^-/\text{\AA}^2$ ) | 110.0                       | Depositor |
| Minimum defocus (nm)                 | 5.0                         | Depositor |
| Maximum defocus (nm)                 | 7.0                         | Depositor |
| Magnification                        | Not provided                |           |
| Image detector                       | GATAN K3 (6k x 4k)          | Depositor |
| Maximum voxel value                  | 105.718                     | Depositor |
| Minimum voxel value                  | -124.889                    | Depositor |
| Average voxel value                  | 8.806                       | Depositor |
| Voxel value standard deviation       | 9.779                       | Depositor |
| Recommended contour level            | Not applicable              |           |
| Tomogram size ( $\text{\AA}$ )       | 13194.24, 12754.433, 5497.6 | wwPDB     |
| Tomogram dimensions                  | 480, 464, 200               | wwPDB     |
| Tomogram angles ( $^\circ$ )         | 90.0, 90.0, 90.0            | wwPDB     |
| Grid spacing ( $\text{\AA}$ )        | 27.488, 27.488, 27.488      | Depositor |

## 2 Tomogram visualisation [i](#)


This section contains visualisations of the EMDB entry EMD-35019. These allow visual inspection of the internal detail of the tomogram and identification of artifacts.

### 2.1 Orthogonal projections [i](#)




The images above show the tomogram projected in three orthogonal directions.

### 2.2 Central slices [i](#)




The images above show central slices of the tomogram in three orthogonal directions.

## 2.3 Largest variance slices [i](#)



The images above show the largest variance slices of the tomogram in three orthogonal directions.

## 2.4 Orthogonal standard-deviation projections (False-color) [i](#)



The images above show the tomogram projected in three orthogonal directions.


## 2.5 Mask visualisation [i](#)

This section was not generated. No masks/segmentation were deposited.

### 3 Tomogram analysis [i](#)

This section contains the results of statistical analysis of the tomogram.

#### 3.1 Voxel-value distribution [i](#)



The voxel-value distribution is plotted in 128 intervals along the x-axis. The y-axis is logarithmic.