



## wwPDB EM Map Validation Summary Report ⓘ

Dec 9, 2020 – 01:38 pm GMT

EMDB ID : EMD-5696  
Title : Electron cryo-microscopy of human cytomegalovirus virion  
Authors : , Dai.XH.; , Yu.XK.; , Gong.H.; , Jiang.XH.; , Abenes.G.; , Liu.HR.; , Shivakoti.S.; , Britt.W.; , Zhu.H.; , Liu.FY.; , Zhou.ZH.  
Deposited on : 2013-06-22  
Resolution : 8.30 Å(reported)

This is a wwPDB EM Map Validation Summary Report for a publicly released PDB entry.

We welcome your comments at [validation@mail.wwpdb.org](mailto:validation@mail.wwpdb.org)

A user guide is available at

<https://www.wwpdb.org/validation/2017/EMMapValidationReportHelp>

with specific help available everywhere you see the ⓘ symbol.

---

The following versions of software and data (see [references ⓘ](#)) were used in the production of this report:

EMDB validation analysis : **FAILED**  
Validation Pipeline (wwPDB-VP) : 2.13

# 1 Experimental information

| Property                             | Value                          | Source    |
|--------------------------------------|--------------------------------|-----------|
| EM reconstruction method             | singleParticle                 | Depositor |
| Imposed symmetry                     | Not Provided                   | Depositor |
| Number of images used                | 1                              | Depositor |
| Resolution determination method      | FSC 0.143 CUT-OFF              | Depositor |
| CTF correction method                | Each particle                  | Depositor |
| Microscope                           | FEI TITAN KRIOS                | Depositor |
| Voltage (kV)                         | 300                            | Depositor |
| Electron dose ( $e^-/\text{\AA}^2$ ) | 25                             | Depositor |
| Minimum defocus (nm)                 | 1.0                            | Depositor |
| Maximum defocus (nm)                 | 3.0                            | Depositor |
| Magnification                        | 59000.0                        | Depositor |
| Image detector                       | GATAN ULTRASCAN 4000 (4k x 4k) | Depositor |